

# Dartmouth Brain Imaging Center (DBIC)


## QA Report 32 CH

11112024

### Measurements


Measurements	Value
mean signal (mean)	1547.1
Signal to Noise Ratio (SNR)	255.8
Signal to Fluctuation Ratio (SFNR)	255.3
Percent Fluctuation	0.06
Drift	0.44
Radius of Decorrelation (RDC)	6.8
Mean Ghost Percentage	2.346
Standard Deviation (std)	0.85

### Signal




## Frequence Spectrum

[mean, SNR, SFNR] = [1547.1 255.8 255.3]




## Raduis of Decorrelation

rdc = 6.8 pixels




## Smoothness - X

Smoothness(FWHM) in mm - X: [min mean max] = [2.061 2.164 2.247]




## Smoothness - Y


Smoothness(FWHM) in mm - Y: [min mean max] = [2.394 2.454 2.516]




## Smoothness - Z




## Center of Mass - X




## Center of Mass - Y




## Center of Mass - Z




## Ghost




## Odd-Even Difference Image



Mean Image



Standard Deviation Image



SFNR Image

