

Dartmouth Brain Imaging Center (DBIC)


QA Report 32 CH

01132025

Measurements


Measurements	Value
mean signal (mean)	1631.4
Signal to Noise Ratio (SNR)	263.6
Signal to Fluctuation Ratio (SFNR)	254.0
Percent Fluctuation	0.06
Drift	0.72
Radius of Decorrelation (RDC)	7.9
Mean Ghost Percentage	1.923
Standard Deviation (std)	0.94

Signal




Frequency Spectrum

[mean, SNR, SFNR] = [1631.4 263.6 254.0]




Radius of Decorrelation

rdc = 7.9 pixels




Smoothness - X

Smoothness(FWHM) in mm - X: [min mean max] = [2.101 2.179 2.280]




Smoothness - Y


Smoothness(FWHM) in mm - Y: [min mean max] = [2.398 2.451 2.519]




Smoothness - Z




Center of Mass - X




Center of Mass - Y




Center of Mass - Z




Ghost




Odd-Even Difference Image



Mean Image



Standard Deviation Image



SFNR Image

