

Dartmouth Brain Imaging Center (DBIC)


QA Report 64 CH

01132025

Measurements


Measurements	Value
mean signal (mean)	2790.7
Signal to Noise Ratio (SNR)	197.3
Signal to Fluctuation Ratio (SFNR)	195.1
Percent Fluctuation	0.06
Drift	0.65
Radius of Decorrelation (RDC)	8.9
Mean Ghost Percentage	2.018
Standard Deviation (std)	1.63

Signal




Frequence Spectrum

[mean, SNR, SFNR] = [2790.7 197.3 195.1]




Raduis of Decorrelation

rdc = 8.9 pixels




Smoothness - X

Smoothness(FWHM) in mm - X: [min mean max] = [2.056 2.219 2.374]




Smoothness - Y

Smoothness(FWHM) in mm - Y: [min mean max] = [2.435 2.551 2.713]




Smoothness - Z

Smoothness(FWHM) in mm - Z: [min mean max] = [1.649 2.172 2.530]




Center of Mass - X


Center of Mass in mm - X: [maxdisplacement drift] = [0.015 0.000]



Center of Mass - Y




Center of Mass - Z




Ghost


Mean of ghost voxels as % of non-ghost [masked] mean
(ghostmean, brightghostmean) = (2.018, 4.673)
(lower is better)




Odd-Even Difference Image



Mean Image



Standard Deviation Image



SFNR Image

